


Disposable medical mask

1、 Finished product drawing:


product implementation standards: YY/T0969-2013

EU standard: CE EN14683:2019

Picture of product outer box: A box of 50 tablets.



Size: Plane lug type 175X95mm



Melt-blown fabric BFE99.2、25X175



2、 Product carton picture specification:

Qty: 50pcs × 40boxes (2000pcs)


Mwas: 48X42.5 X50.3cm

(G. W.) : 9.5kg (N.W.) : 8.0kg

Validity: 2years

Quality guarantee period: Two years

date of manufacture : See product code printing



3. Company qualification

Registration certificate of medical device B registration certificate of disposable medical mask C production license of medical device


中华人民共和国医疗器械注册证

注册证编号：浙械注准20202141085

注册人名称	杭州天诚药业有限公司
注册人住所	浙江省萧山经济技术开发区桥南区块春江路1号
生产地址	浙江省萧山经济技术开发区桥南区块春江路1号
代理人名称	不适用
代理人住所	不适用
产品名称	一次性使用医用口罩
型号、规格	型号：非无菌型耳挂式；规格：175mm×95mm、170mm×100mm、145mm×95mm、140mm×90mm、125mm×90mm。
结构及组成	产品由口罩体、鼻夹和口罩带组成。
适用范围	用于普通环境下的一次性卫生护理，不作外科手术时特殊防护用。
附件	本产品执行YY/T 0969-2013《一次性使用医用口罩》标准要求（除4.7.2 无菌、4.8 环氧乙烷残留量、4.9 生物学评价外）。
其他内容	/
备注	本产品为防控新型冠状病毒肺炎疫情的应急审批产品，注册证有效期为6个月，产品标签和说明书上应醒目标注“仅供防控疫情应急使用”。

审批部门：浙江省药品监督管理局

批准有效
2020年03月19日



营业执照

(副-社章) 统一社会信用代码 91330109609128303D (1/1)

名称 杭州天诚药业有限公司
类型 有限责任公司
住所 浙江省萧山经济技术开发区桥南区块春江路1号
法定代表人 潘观根
注册资本 壹仟玖佰肆拾万元整
成立日期 1993年05月22日
营业期限 1993年05月22日至2058年05月21日
经营范围 生产：医疗器械（具体范围见《医疗器械生产企业许可证》，保健食品（具体范围见《保健食品生产企业许可证》）；卫生用品；抗（抑）菌剂（液体）（净化）（以上经营范围在许可证有效期内方可经营）
** 销售：化妆品；收购：本企业所需的原料材料（限直接向第一产业的原始生产者收购）；生物技术研发；医药信息咨询；企业营销策划；房屋租赁、物业服务**（依法须经批准的项目，经相关部门批准后方可开展经营活动）


多证合一 登记机关

2017年10月10日

企业应当于每年1月1日至6月30日通过浙江省企业信用信息公示系统报送年度报告

<http://gsxt.zjmc.gov.cn/>

企业信用信息公示系统网址： 中华人民共和国国家市场监督管理总局监制



医疗器械生产许可证

许可证编号：浙食药监械生产许20110094号


企业名称：杭州天诚药业有限公司
生产地址：浙江省萧山经济技术开发区桥南区块春江路1号

法定代表人：潘观根
生产范围：旧版：第二类 6801-2-一次性使用无菌手术刀、6856-1-供气系统、6864-0-其他、6864-2-敷料、护创材料，第三类 6864-2-敷料、护创材料；新版：第二类 14-14-医护人员防护用品***

企业负责人：潘国英

住所：浙江省萧山经济技术开发区桥南区块春江路1号 发证部门：浙江省药品监督管理局


有效期限：至 2021年7月4日 发证日期：2020年3月19日



4. Instructions for export of disposable medical masks (non sterile) to EU

EU medical / non-medical masks need to be attached with CE mark, but the applicable standards are different. In the EU, medical masks belong to class I devices, which are divided into class I non sterile and sterile. CE mark shall be attached according to EU medical device directive 93 / 42 / EEC (MDD) or EU medical device regulation eu2017 / 745 (MDR). The corresponding standard is en14683. The outer packaging or test report and certificate of masks exported to the EU Those with the above contents can be determined as medical masks. Non sterile masks need CE self-compliance declaration, and do not need to pass the certification of the notified body. After preparing the corresponding documents, test reports and other data, the declaration of conformity can be completed by yourself.

EC DECLARATION OF CONFORMITY

MANUFACTURER	HangZhou TianCheng Pharmaceutical Co., Ltd.
Adress	Chunjiang Road No1, Qiaonan Zone, Xiaoshan Economic & Technological Development Zone, Zhejiang Province, China
The undersigned Company certifies under its sole responsibility that the item of equipment specified below satisfies the requirements of the Medical Devices 93/42/EEC which is apply to it. The item of equipment identified below has been subject to internal manufacturing checks with monitoring of the final assessment by HangZhou TianCheng Pharmaceutical Co., Ltd.	
PRODUCT	Disposable surgical mask
MODEL / TYPE	175mm×95mm/Type II
DIRECTIVES	
Medical Devices 93/42/EEC	
Regulations Applied acc. to HARMONIZE STANDARDS	
EN14683:2019 Type II	
Place and date of issue	: HangZhou City, Zhejiang Province, China /2020.04.02
Name and position of authorized person	: Pan Guoying /General Manager
Singanature of authorized person	: Pan Guange 

Technical Construction File EN 14683:2019 Medical face masks - Requirements and test methods	
Report reference No.:	TMZJ20031022146
Completed by (+ signature):	Stephen Zhang / Test Engineer
Approved by (+ signature):	Kosco Vant / Project Manager
Date of issue:	March 10, 2020
Reviewing laboratory:	Shanghai Global Testing Services Co., Ltd.
Reviewing location:	Floor 2nd, Building D-1, No. 126, Shenfu Road, Minhang District, Shanghai, China.
Applicant:	HangZhou TianCheng Pharmaceutical Co., Ltd.
Address:	Chunjiang Road No1, Qiaonao Zone, Xiaoshan Economic & Technological Development Zone, Zhejiang Province
Manufacturer:	HangZhou TianCheng Pharmaceutical Co., Ltd.
Address:	Chunjiang Road No1, Qiaonao Zone, Xiaoshan Economic & Technological Development Zone, Zhejiang Province
Factory:	The same as Manufacturer
Address:	The same as Manufacturer
Standard:	EN 14683:2019
Review Report Form No.:	14683
TRF originator:	GTS
Master TRF:	Reference No. EN 14683:2019
Review procedure:	GTS
Type of Review object:	Disposable Surgical Mask
Trademark:	-
Model/type reference:	175mm x 95mm, 170mm x 100mm, 145mm x 95mm, 140mm x 90mm, 125mm x 90mm
Rating:	175mm x 95mm, 170mm x 100mm, 145mm x 95mm, 140mm x 90mm, 125mm x 90mm



Possible review case verdicts: - review case does not apply to the test object.....: N(A) - review object does meet the requirement.....: P(all) - review object does not meet the requirement.....: F(all)	
General remarks: [see remark #] refers to a remark appended to the report. [see appended table] refers to a table appended to the report. Throughout this report a comma is used as the decimal separator. The review results presented in this report relate only to the object reviewed. This report shall not be reproduced except in full without the written approval of the third party.	
Testing: Date of receipt of review item:	March 08,2020
Date(s) of performance of review:	March 08,2020 to March 10,2020
General product information: Disposable Surgical Mask	
Summary of reviewing: This review report includes: Annex I: 1 page(s) of photo documentation.	
Copy of marking plate	
Disposable Surgical Mask, Model 175mm x 95mm, 170mm x 100mm, 145mm x 95mm, 140mm x 90mm, 125mm x 90mm	Marking
HangZhou TianCheng Pharmaceutical Co., Ltd.	


4	Classification	Type II R	P
	Medical face masks specified in this European Standard are classified into two types (Type I and Type II) according to bacterial filtration efficiency whereby Type II is further divided according to whether or not the mask is splash resistant. The 'S' signifies splash resistance.		
5	Requirements		
5.1	General		
5.1.1	Materials and construction		
	The medical face mask is a medical device, generally composed of a filter layer that is placed, bonded or moulded between layers of fabric. The medical face mask shall not disintegrate, split or tear during intended use. In the selection of the filter and layer materials, attention shall be paid to cleanliness.		P
5.1.2	Design		
	The medical face mask shall have a means by which it can be fitted closely over the nose, mouth and chin of the wearer and which ensures that the mask fits closely at the sides. Medical face masks may have different shapes and constructions as well as additional features such as a face shield (to protect the wearer against splashes and droplets) with or without anti-fog function, or a nose bridge (to enhance fit by conforming to the nose contours).		P
5.2	Performance requirements		
5.2.1	General		
	All tests shall be carried out on finished products or samples cut from finished products.		P
5.2.2	Bacterial filtration efficiency (BFE)	BFE ≥ 98%	P
	When tested in accordance with Annex B, the BFE of the medical face mask shall conform to the minimum value given for the relevant type in Table 1. For thick and rigid masks such as rigid ductal or cup masks the test method may not be suitable as a proper seal cannot be maintained in the cascade impactor. In these cases, another valid equivalent method shall be used to determine the BFE. When a mask consists of two or more areas with different characteristics or different layer-composition, each panel or area shall be tested individually. The lowest performing panel or area shall determine the BFE value of the complete mask.		
5.2.3	Breathability		
	When tested in accordance with Annex C, the		P

	differential pressure of the medical face mask shall conform to the value given for the relevant type in Table 1. If the use of a respiratory protective device as face mask is required in an operating theatre and/or other medical settings, it might not fulfil the performance requirements with regard to differential pressure as defined in this European Standard. In such case, the device should fulfil the requirement as specified in the relevant Personal Protective Equipment (PPE) standard(s).																																
5.2.4	Splash resistance		N/A																														
	When tested according to EN ISO 22909:2004 the resistance of the medical face mask to penetration of sprays of liquid shall conform to the minimum value given for Type II(S) in Table 1.																																
5.2.5	Microbial cleanliness (bioburden)		P																														
	When tested according to EN ISO 11737-1:2018 the bioburden of the medical mask shall be <= 30 CFU/g tested (see Table 1). To determine the mask's bioburden according to EN ISO 11737-1:2018, refer to the procedure as described in Annex D. The number of masks that shall be tested is minimum 5 of the same batch/lot. Other test conditions as described in EN ISO 11737-1:2018 may be applied. In the test report, indicate the total bioburden per individual mask and based on the mask weight, the total bioburden per gram.																																
5.2.6	Biocompatibility		P																														
	According to the definition and classification in EN ISO 10993-1:2009, a medical face mask is a surface device with limited contact. The manufacturer shall complete the evaluation of the medical face mask according to EN ISO 10993-1:2009 and determine the applicable toxicology testing regime. The results of testing should be documented according to the applicable parts of the EN ISO 10993 series. The test results shall be available upon request.																																
5.2.7	Summary of performance requirements																																
	Table 1 - Performance requirements for medical face masks																																
	<table border="1"> <thead> <tr> <th>Type</th> <th>Filter</th> <th>Type I</th> <th>Type II</th> <th>Type II(S)</th> </tr> </thead> <tbody> <tr> <td>Bacterial Filtration Efficiency (BFE)</td> <td>≥ 95%</td> <td>≥ 95%</td> <td>≥ 98%</td> <td>≥ 98%</td> </tr> <tr> <td>Bioburden</td> <td><= 30 CFU/g</td> <td><= 30 CFU/g</td> <td><= 30 CFU/g</td> <td><= 30 CFU/g</td> </tr> <tr> <td>Differential Pressure</td> <td><= 150 Pa</td> <td><= 150 Pa</td> <td><= 150 Pa</td> <td><= 150 Pa</td> </tr> <tr> <td>Resistance to Penetration of Liquid</td> <td><= 100 Pa</td> <td><= 100 Pa</td> <td><= 100 Pa</td> <td><= 100 Pa</td> </tr> <tr> <td>Biocompatibility</td> <td>EN ISO 10993-1:2009</td> <td>EN ISO 10993-1:2009</td> <td>EN ISO 10993-1:2009</td> <td>EN ISO 10993-1:2009</td> </tr> </tbody> </table>	Type	Filter	Type I	Type II	Type II(S)	Bacterial Filtration Efficiency (BFE)	≥ 95%	≥ 95%	≥ 98%	≥ 98%	Bioburden	<= 30 CFU/g	<= 30 CFU/g	<= 30 CFU/g	<= 30 CFU/g	Differential Pressure	<= 150 Pa	<= 150 Pa	<= 150 Pa	<= 150 Pa	Resistance to Penetration of Liquid	<= 100 Pa	<= 100 Pa	<= 100 Pa	<= 100 Pa	Biocompatibility	EN ISO 10993-1:2009	EN ISO 10993-1:2009	EN ISO 10993-1:2009	EN ISO 10993-1:2009		
Type	Filter	Type I	Type II	Type II(S)																													
Bacterial Filtration Efficiency (BFE)	≥ 95%	≥ 95%	≥ 98%	≥ 98%																													
Bioburden	<= 30 CFU/g	<= 30 CFU/g	<= 30 CFU/g	<= 30 CFU/g																													
Differential Pressure	<= 150 Pa	<= 150 Pa	<= 150 Pa	<= 150 Pa																													
Resistance to Penetration of Liquid	<= 100 Pa	<= 100 Pa	<= 100 Pa	<= 100 Pa																													
Biocompatibility	EN ISO 10993-1:2009	EN ISO 10993-1:2009	EN ISO 10993-1:2009	EN ISO 10993-1:2009																													

6	Marking, labelling and packaging		P
	Annex I, § 13, of the Medical Devices Directive (93/42/EEC) or Annex I, § 23, of the Medical Device Regulation (EU) 2017/745 specifies the information that should be specified on the packaging in which the medical face mask is supplied. The following information shall be supplied: a) number of this European Standard; b) type of mask (as indicated in Table 1); c) EN ISO 15223-1:2016 and EN 1041:2009+A1:2013 should be considered.		

- End of Review Report -

Type of equipment, model: Disposable Surgical Mask
175mm x 95mm, 170mm x 100mm, 145mm x 95mm, 140mm x 90mm, 125mm x 90mm



- End of Annex I -

5、 Disposable medical mask test report



检验报告

Test Report

报告编号: Z20200669

产品名称: 一次性使用医用口罩

规格型号: 型号: 非灭菌型, 规格: 175mm×95mm

委托单位: 杭州天诚药业有限公司

检验类别: 委托检验

浙江省医疗器械检验研究院
ZHEJIANG INSTITUTE OF MEDICAL DEVICE SUPERVISION AND TESTING

浙江省医疗器械检验研究院

检验报告

报告编号: Z20200669	共 6 页 第 1 页		
样品名称	一次性使用医用口罩		
规格型号	型号: 非灭菌型, 规格: 175mm×95mm	检验类别	委托检验
委托人/单位	杭州天诚药业有限公司		
受托单位名称	杭州天诚药业有限公司		
制造单位名称	杭州天诚药业有限公司		
取样方式	送样	抽样地点	/
抽样日期	/	抽样基数	/
抽样人	/	样品接受日期	2020-03-09
样品数量	12包 (共60只)	样品生产日期	20200309
样品批号/编号	200301		
检验依据	YY/T0969-2013		
检验项目	全性能检验(4.7.2, 4.8, 4.9除外)		
检验日期	2020年03月09日~2020年03月23日		
检验结论	被检样品所检项目符合YY/T 0969-2013《一次性使用医用口罩》的要求。		

批准: 张华 审核: 周明 主检: 王磊群
职务: 授权签字人 日期: 2020.03.24 日期: 2020.03.24 日期: 2020.03.24

检验报告

报告编号: Z20200669	共 6 页 第 2 页		
检验地点	省质、宁波实验室	本报告共有3张照片	
分包检验项目	/		
委托人/单位资料	电话: 0571-82697225	邮政编码	/
	地址: 浙江省萧山经济技术开发区桥西区块春江路1号		
采用的检验方法	<input checked="" type="checkbox"/> 标准方法; <input type="checkbox"/> 客户实物的方法; <input type="checkbox"/> 本实验室提供的方法; <input type="checkbox"/> 其他方法;		
评估测量不确定度的声明/信息	/		
意见和解释	/		
样品描述	/样品外观描述: a) 结构组成:/; b) 原理、用途:/; c) 主要特征参数:外观、结构与尺寸、热封、口罩带、细菌过滤效率、透气阻力、微生物限度; d) 样品状态:完好。		
其他说明	/		

检测结果汇总

报告编号: Z20200669	共 6 页 第 3 页				
序号	检测项目	标准条款	标准要求	实测结果	判定结论
1	外观	4.1	口罩外观应整洁、形状完好,表面不得有破损、污渍	符合	合格
2	结构与尺寸	4.2	应符合设计尺寸,最大偏差应不超过±5%	1 2 3	合格
			口罩长度: 175mm	174 175 174	
			口罩宽度: 95mm	94 94 94	
3	鼻夹	4.3.1	口罩上应配有鼻夹,鼻夹应由弹性材料制成	符合	合格
		4.3.2	鼻夹长度应不小于8.0cm	1 2 3 10.1 9.8 10.1	合格
4	口罩带	4.4.1	口罩带应佩戴方便	符合	合格
		4.4.2	绳状口罩带与口罩体连接点的耐牵拉力应不小于10N	符合	合格
5	细菌过滤效率(BFE)	4.5	口罩的细菌过滤效率应不小于95%	1 2 3 99.5% 99.6% 99.7%	合格
6	通气阻力	4.6	口罩两侧面进行气体交换的静气阻力应不大于49Pa/cm ²	1 2 3 27 29 32	合格
7	微生物限度	4.7.1	细菌菌落总数: ≤100CFU/g	28CFU/g	合格
			大肠菌群: 不得检出	未检出	合格

检测结果汇总

报告编号: Z20200609

共 6 页 第 4 页

序号	检测项目	标准条款	标准要求	实测结果	判定结论
			绿脓杆菌; 不得检出	未检出	合格
			金黄色葡萄球菌; 不得检出	未检出	合格
			溶血性链球菌; 不得检出	未检出	合格
			真菌; 不得检出	未检出	合格

本页以下空白

报告编号: Z20200609

共 6 页 第 5 页



报告编号: Z20200609

共 6 页 第 6 页



产品名称: 一次性使用医用口罩
 型号: 非灭菌型
 规格: 175mm×95mm 包装: 5个/包
 批号: 200301
 生产日期: 2020年03月09日
 生产单位: 杭州天诚药业有限公司

Bacterial filtration rate (BFE) test results: 99.2%

6、 Factory production picture



7、 Cooperation terms (RMB settlement)

50% of the total payment shall be paid on the day of signing the contract. Both parties shall agree on the capacity delivery form and take the agreed delivery quantity on the same day

50% of the balance before delivery

According to the announcement of the State Drug Administration of the General Administration of Customs of the Ministry of Commerce of the people's Republic of China on the orderly export of medical materials (No. 5 in 2020), it is clear that:

Our company's products: disposable medical masks are among them.



中华人民共和国商务部
MINISTRY OF COMMERCE OF THE PEOPLE'S REPUBLIC OF CHINA

首页
机构设置
政务公开
政务服务
互动交流
公共服务

> 首页 > 政务公开 > 对外经贸管理

商务部 海关总署 国家药品监督管理局公告2020年第5号 关于有序开展医疗物资出口的公告

文章来源：商务部对外公告 2020-03-31 22:47 文章来源：原创 作者：佚名 英文：【[查看详情](#)】

当前，全球疫情呈加速扩散蔓延态势，在做好自身疫情防控的基础上，有序开展医疗物资出口是深化疫情防控国际合作、共同应对全球公共卫生危机的关键举措。在疫情防控特殊时期，为有效支持全球抗击疫情，保证产品质量安全、规范出口秩序，自4月1日起，出口新型冠状病毒疫情防控物资，医用口罩、医用防护服、呼吸机、红外体温计的企业向海关报关时，须提供书面或电子声明（模板见附件1），承诺出口产品已取得我国医疗器械产品注册证书（相关注册信息见附件2），符合进口国（地区）的质量标准要求。海关凭药品监督管理部门批准的医疗器械产品注册证书验放。上述医疗物资出口质量监管措施将视疫情发展动态调整。

有关医疗物资出口企业要确保产品质量安全，符合相关标准要求，积极支持国际社会共同抗击疫情。

附件：1. 出口医疗物资声明模板.doc

2. 我国相关医疗器械产品注册信息（国家药监局网站www.cma.gov.cn动态更新）.xlsx

商务部 海关总署 国家药品监督管理局
2020年3月31日


675	浙械注准20202141048	一次性使用医用口罩	杭州山友医疗器械有限公司
676	浙械注准20202141050	普通医用口罩	绍兴瑞凯防护用品有限公司
677	浙械注准20202141051	一次性使用医用口罩	浙江科瑞医药科技有限公司
678	浙械注准20202141052	一次性使用医用口罩	浙江奕达皮革服饰有限公司
679	浙械注准20202141053	一次性使用医用口罩	浙江佰意智造服饰有限公司
680	浙械注准20202141061	一次性使用医用口罩	宏昌生物医疗科技(平湖)有限公司
681	浙械注准20202141065	一次性使用医用口罩	浙江银达生物技术有限公司
682	浙械注准20202141066	一次性使用医用口罩	宁波健翔包装材料科技有限公司
683	浙械注准20202141073	一次性使用医用口罩	浙江天使医疗器械有限公司
684	浙械注准20202141074	一次性使用医用口罩	杭州可靠护理用品股份有限公司
685	浙械注准20202141076	一次性使用医用口罩	宁波金格奥电器股份有限公司
686	浙械注准20202141078	一次性使用医用口罩	浙江家壹卫生用品有限公司
687	浙械注准20202141079	一次性使用医用口罩	浙江龙德医药有限公司
688	浙械注准20202141080	一次性使用医用口罩	杭州立山皮件有限公司
689	浙械注准20202141082	一次性使用医用口罩	杭州万事利丝绸数码印花有限公司
690	浙械注准20202141085	一次性使用医用口罩	杭州天诚药业有限公司
691	浙械注准20202141086	一次性使用医用口罩	拓胜防护用品(宁波)有限公司
692	浙械注准20202141087	一次性使用医用口罩	浙江隆泰医疗科技股份有限公司
693	浙械注准20202141090	一次性使用医用口罩	金华德邦医疗器械有限公司
694	浙械注准20202141091	一次性使用医用口罩	宁波欧涵医疗器械有限公司
695	浙械注准20202141092	一次性使用医用口罩	浙江文明德生物科技有限公司
696	浙械注准20202141094	一次性使用医用口罩	绍兴港峰用品有限公司
697	浙械注准20202141096	一次性使用医用口罩	永康市巨立防护用品有限公司
698	浙械注准20202141099	一次性使用医用口罩	杭州永利百合医疗器械有限公司
699	浙械注准20202141101	一次性使用医用口罩	浙江蓝天制衣有限公司
700	浙械注准20202141103	一次性使用医用口罩	德普斯医疗器械湖州有限公司
701	浙械注准20202141105	一次性使用医用口罩	浙江蓝天鹤舞控股有限公司
702	浙械注准20202141106	一次性使用医用口罩	浙江建安检测研究院有限公司
703	浙械注准20202141108	一次性使用医用口罩	浙江京环医疗用品有限公司
704	浙械注准20202141110	一次性使用医用口罩	浙江欣富无纺布科技有限公司
705	浙械注准20202141112	一次性使用医用口罩	杭州润恒医疗器械有限公司
706	浙械注准20202141113	一次性使用医用口罩	嘉兴市正群医疗器械有限公司
707	浙械注准20202141117	一次性使用医用口罩	杭州达维先医药科技有限公司
708	浙械注准20202141118	一次性使用医用口罩	浙江欧洁科技股份有限公司
709	浙械注准20202141119	一次性使用医用口罩	江山市美绮服饰有限公司
710	浙械注准20202141122	一次性使用医用口罩	金华家大夫医护用品有限公司
711	浙械注准20202141124	一次性使用医用口罩	德普斯医疗器械湖州有限公司
712	浙械注准20202141125	一次性使用医用口罩	浙江华光胶囊股份有限公司
713	浙械注准20202141126	一次性使用医用口罩	浙江康柏斯医疗科技有限公司
714	浙械注准20202141127	一次性使用医用口罩	冀发集团有限公司
715	津械注准20202140842	一次性使用医用口罩	天津骏发森达卫生用品有限公司

新冠病毒检测试剂
呼吸机
医用防护服
医用防护口罩
医用外科口罩
一次性使用医用口罩
红外体温计

Intco Medical disposable Nitrile gloves



Without powder
Without smell
Both hand use
Oil proof
Waterproof





Intco Medical disposable PVC gloves



Without powder
Without smell
Both hand use
Oil proof
Waterproof



Package information:

Products	Size	N.w.(kgs)	G.w.(kgs)	Package	Package size	Pictures
Nitrile Gloves	S	3.0	4.0	100pcs/box 10 boxes/ package	Outside :33.5*26.5*26.5CM	
	M	3.5	4.5		Inside :24*12.5*6.3CM	
	L	4.0	5.0		1300 Package/20ft FCL	
	XL	4.5	5.5		3300 Package/40ft FCL	
PVC Gloves	S	4.0	5.0	100pcs/box 10 boxes/ package	Outside :33.5*26.5*26.5CM	
	M	4.5	5.5		Inside :24*12.5*6.3CM	
	L	5.0	6.0		1300 Package/20ft FCL	
	XL	5.5	6.5		3300 Package/40ft FCL	



SGS

TEST REPORT

No. : QZIN1510045067MR
Date : Oct 30, 2015
Page: 2 of 6

Summary of Results:

No.	Test Item	Test Method	Result	Conclusion	
1	Length	ISO 4592:1992	Sample 1	244 mm	/
			Sample 2	241 mm	
			Sample 3	239 mm	
2	Tensile Strength	With reference to ASTM D412-06a(2013) Method A and client's requirement	Sample 1	11.6 MPa	/
			Sample 2	17.5 MPa	
			Sample 3	16.4 MPa	
3	Thickness	ISO 4593:1993	Sample 1	0.067 mm	/
			Sample 2	0.066 mm	
			Sample 3	0.066 mm	

Note: Pass : Meet the requirements;
Fail : Does not meet the requirements;
/ : Not Apply to the judgment.

This document is issued by the Company subject to its General Conditions of Purchase and/or general conditions of sale or availability of the Company's products. The Company shall not be held liable for any loss or damage caused by the use of the information contained herein. The Company shall not be held liable for any loss or damage caused by the use of the information contained herein. The Company shall not be held liable for any loss or damage caused by the use of the information contained herein. The Company shall not be held liable for any loss or damage caused by the use of the information contained herein.

Test Report No. 7191203047-EEC19-WBH
dated 28 Jan 2019

TUV
PSB Singapore

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, the report is governed by the terms set out within this report.

SUBJECT:
Testing of Disposable Vinyl Gloves (PVC) submitted by Shandong Inteco Medical Products Co., Ltd. on 16 Jan 2019.

TESTED FOR:
Shandong Inteco Medical Products Co., Ltd.
No 9688 Qiawang Road, Naoshan Industry Park, Qingzhou, Shandong, China

TEST DATE:
18 Jan 2019

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Disposable Vinyl Glove (PVC)	Clear	/	M	200	Shandong Inteco Medical Products Co., Ltd.

Lot size as specified by client: 35,001 to 150,000 pieces

METHOD OF TEST:
EN 455-1:2000 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes

TUV
PSB

Laboratory:
TUV SÜD PSB Pte. Ltd.
No 1 Science Park Drive
Singapore 118221



Issued to:

Shandong IntcoMedical Products Co. Ltd
Qiwang Road, Naoshan Industrial Park
Qingzhou
Shandong
262506
China

Notified Body: 2777

SATRA customer number: P1720

EU Type-Examination Certificate

Certificate number: 2777/11804-01/E00-00

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:

Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

Product reference:

Description:

697024575

Five fingered disposable nitrile non-sterile gloves.

Blue 697024575 601-605

Violet 697024575 631-635

White 697024575 641-645

Black 697024575 651-655

Sizes:

6/XS, 6.5/S, 7/M, 8/L, 9/XL

Classification:

EN ISO 374-1:2016+A1:2018 /Type B

40% Sodium Hydroxide (K)

30% Hydrogen peroxide (P)

37% Formaldehyde (T)

EN ISO 374-5: 2016

Protection against Bacteria and fungi

Protection against viruses

Level

6

2

3

Pass

Pass

EN 374-4: 2013

-11.5 %

-9.5%

7.4 %

Standards/Technical specifications applied:

EN 420: 2003+A1: 2009; EN ISO 374-1:2016+A1:2018; EN ISO 374-5:2016

Technical reports/Approval documents:

SATRA: CHT0278438/1848

SGS: QDHL18060131130T, CH:TX:6420074520, CH:TX: 9420020333, CH:TX: 9420029243 CH:TX: 9420026599-1, CH:TX:

9420014953-1, CH:TX: 9420026316-1, CH:TX: 9420614959

TUV: 721642857-2

Signed on behalf of SATRA:

Tara Saunders

Geoff Graham

Date first issued: 30/01/2019

Date of issue: 30/01/2019

Expiry date: 30/01/2024